



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME TSH

Code	86516		
Lot No.	001	2025-01	2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D3290	TSH	001	2025-01	6X6
B3290	CALIBRATOR TSH	001	2025-01	1
C3290	CONTROL SERUM TSH	001	2025-01	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	389	≥ 100
Control Serum (uIU/ml)	2,4	1,5-4,7

CALIBRATOR

9	1	9	0	0	1	6	7	1	4	1	9	5	5	4	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB



CONTROL SERUM

9	1	9	0	0	1	6	1	3	2	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 20/02/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.