



QUALITY CONTROL CERTIFICATE

KIT NAME HELICOBACTER PYLORI IgG -CHORUS

Code	81060		
Lot No.	004	2025-01	2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D2391	Helicobacter Pylori IgG	004	2025-01	6X6
B2392	CALIBRATORE Helicobacter Pylori IgG	004	2025-01	1
C2391	CONTROL SERUM Helicobacter Pylori IgG	004	2025-01	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	387	≥ 100
Control Serum (UA/mL)	110,6	67,9-200,0
Negative Control (UA/mL)	<5,0	< 5

CALIBRATOR

0	0	8	0	0	4	6	1	2	0	3	4	2	9	4	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB

008004612034294600  
Helicobacter pylori IgG  
CALIBRATOR 0.425 mL IVD  
LOT 004  
2025-01  
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

0	0	8	0	0	4	6	2	5	4	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB

008004625460000001  
Helicobacter pylori IgG  
CONTROL 0.425 mL IVD  
LOT 004  
2025-01  
DIESSSE Monteriggioni Italy CE

**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents Luca

DATE 20/02/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.