

QUALITY CONTROL CERTIFICATE

KIT NAME DEAMIDATED GLIADIN-A - CHORUS

Code	86098		
Lot No.	008	2025-02	2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D8650	Deamidated Gliadin-A	008	2025-02	6X6
B8650	CALIBRATORE Deamidated Gliadin-A	008	2025-02	1
C8650	CONTROL SERUM Deamidated Gliadin-A	008	2025-02	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	522	≥ 100
Control Serum (AU/mL)	79,2	50,0-233,3

CALIBRATOR

2	5	4	0	0	8	6	3	2	0	3	0	3	5	4	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature FB



CONTROL SERUM

2	5	4	0	0	8	6	2	0	3	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. FB

DATE 23.02.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.