



QUALITY CONTROL CERTIFICATE

KIT NAME DEAMIDATED GLIADIN-G - CHORUS

Code	86100		
Lot No.	008	2025-02	2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D8660	Deamidated Gliadin-G	008	2025-02	6X6
B8660	CALIBRATORE Deamidated Gliadin-G	008	2025-02	1
C8660	CONTROL SERUM Deamidated Gliadin-G	008	2025-02	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	679	≥ 100
Control Serum (AU/mL)	49,4	26,7-72,7

CALIBRATOR

2	5	5	0	0	8	6	1	1	9	3	0	5	5	5	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature FB



CONTROL SERUM

2	5	5	0	0	8	6	1	3	2	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 26/02/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.