

DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME FSH

Code	86510		
Lot No.	001	2024-10	2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D3230	FSH	001	2024-10	6X6
B3230	CALIBRATOR FSH	001	2024-11	1
C3230	CONTROL SERUM FSH	001	2024-11	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	260	≥ 100
Control Serum (mIU/ml)	11,3	7,6-14,2

CALIBRATOR

9	2	0	0	0	1	5	4	1	5	3	7	3	7	4	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *MB*

920001541537374900
FSH
CALIBRATOR 0,600 mL IVD
LOT 001
2024-11 2°C-8°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

9	2	0	0	0	1	5	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *MB*

920001519350000001
FSH
CONTROL+ 1 mL IVD
LOT 001
2024-11 2°C-8°C
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE *23.02.2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.