

QUALITY CONTROL CERTIFICATE

KIT NAME SARS-CoV-2 "Neutralizing" Ab - CHORUS

Code	81408		
Lot No.	002	2025-01	2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D4240	SARS-CoV-2 "Neutralizing" Ab	002	2025-01	6X6
B4240	CALIBRATORE SARS-CoV-2 "Neutralizing" Ab	002	2025-02	1
C4240	CONTROL SERUM SARS-CoV-2 "Neutralizing" Ab	002	2025-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	446	≥ 100
Control Serum (BAU/mL)	171,7	140,0-221,9

CALIBRATOR

1	7	0	0	0	2	6	2	2	0	3	1	7	9	3	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature FB



CONTROL SERUM

1	7	0	0	0	2	6	0	9	1	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... FB

DATE 28/02/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.