



QUALITY CONTROL CERTIFICATE

KIT NAME MYCOPLASMA PNEUMONIAE IgM - CHORUS

Code	81035		
Lot No.	007	2025-05	2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D2261	MYCOPLASMA PNEUMONIAE IgM	007	2025-05	6x6
B2261	CALIBRATORE MYCOPLASMA PNEUMONIAE IgM	007	2025-05	1
C2261	CONTROL SERUM MYCOPLASMA PNEUMONIAE IgM	007	2025-05	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	595	≥ 100
Control Serum (INDEX)	1,6	1,2-2,2
Negative Control (INDEX)	0,1	≤ 0,7

CALIBRATOR

0	6	7	0	0	7	7	0	0	0	0	0	0	6	1	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

067007700000610000  
Mycoplasma pneumoniae IgM  
CALIBRATOR 0.175 mL [IVD]  
LOT 007  
2025-05 2°C-8°C B2261  
DIESSE Monteriggioni Italy CE

CONTROL SERUM

0	6	7	0	0	7	7	1	2	2	2	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

0670077122200000001  
Mycoplasma pneumoniae IgM  
CONTROL+ 0.425 mL [IVD]  
LOT 007  
2025-05 2°C-8°C C2261  
DIESSE Monteriggioni Italy CE

**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

DATE 28.02.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.