



QUALITY CONTROL CERTIFICATE

KIT NAME 88000-D70 ACARUS SIRO

Code	88000-D70		
Lot No.	002	2025-01	2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D1922	88000-D70 ACARUS SIRO	002	2025-01	2x6
B1921	CALIBRATORE 88000-D70 ACARUS SIRO	002	2025-02	1
C1921	CONTROL SERUM 88000-D2 ACARUS SIRO	002	2025-02	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	284	≥ 100
Control Serum (KU/L)	8,2	5,8-14,3

CALIBRATOR

6	5	5	0	0	2	6	5	2	1	6	0	2	6	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB

655002652160268100  
Acarus siro  
CALIBRATOR 0.250 mL IVD  
LOT 002  
2025-02 2°C-8°C  
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

6	5	5	0	0	2	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FO

655002627500000001  
Acarus siro  
CONTROL + 0.450 mL IVD  
LOT 002  
2025-02 2°C-8°C  
DIESSSE Monteriggioni Italy CE

**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]  
DATE 07/03/2024

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**  
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.