

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

**QUALITY CONTROL CERTIFICATE**

**KIT NAME 88000-F14 SOY BEAN**

Code	88000-F14		
Lot No.	002	2025-01	2024-03

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D9831	88000-F14 SOY BEAN	002	2025-01	2x6
B9830	CALIBRATORE 88000-F14 SOY BEAN	002	2025-02	1
C9830	CONTROL SERUM 88000-F14 SOY BEAN	002	2025-02	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	284	≥ 100
Control Serum (KU/L)	8,2	5,8-14,3

**CALIBRATOR**

6	3	0	0	0	2	6	5	2	1	6	0	2	6	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*

630002652160268100  
Soy bean  
CALIBRATOR 0,250 mL IVD  
LOT 002  
2025-02 2°C-8°C  
DIESSE Monteriggioni Italy CE

**CONTROL SERUM**

6	3	0	0	0	2	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*

630002627500000001  
Soy bean  
CONTROL 0,450 mL IVD  
LOT 002  
2025-02 2°C-8°C  
DIESSE Monteriggioni Italy CE

**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents... *[Signature]*  
DATE *07.03.2024*

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**  
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.