

QUALITY CONTROL CERTIFICATE

KIT NAME dsDNA-G - CHORUS

Code	86032			
Lot No.	011	2025-02	2024-02	

REAGENTS

Label	Description	Lot.	Exp.	NO
D3281	dsDNA-G	011	2025-02	6X6
B3281	CALIBRATORE dsDNA-G	011	2025-02	1
C3281	CONTROL SERUM dsDNA-G	011	2025-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	488	≥ 100
Control Serum (IU/mL)	40,6	33,9-66,7

CALIBRATOR

2	5	1	0	1	1	6	5	2	4	3	3	4	8	3	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

251011652433483600
dsDNA-G
CALIBRATOR 0.176 mL IVD
LOT 011
2025-02 2°C-8°C
DIESE Monteriggioni Italy CE

CONTROL SERUM

2	5	1	0	1	1	6	1	5	2	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

251011615270000001
dsDNA-G
CONTROL+ 0.425 mL IVD
LOT 011
2025-02 2°C-8°C
DIESE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. [Signature]

DATE 07/03/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.