



QUALITY CONTROL CERTIFICATE

KIT NAME tTg-G - CHORUS

| | | | |
|---------|-------|---------|---------|
| Code | 86060 | | |
| Lot No. | 004 | 2025-01 | 2024-01 |

REAGENTS

| Label | Description | Lot. | Exp. | NO |
|-------|---------------------|------|---------|-----|
| D3451 | tTg-G | 004 | 2025-01 | 6X6 |
| B3451 | CALIBRATORE tTg-G | 004 | 2025-01 | 1 |
| C3451 | CONTROL SERUM tTg-G | 004 | 2025-01 | 1 |

CONTROLS

| | Found results | Expected results |
|-----------------------|---------------|------------------|
| Calibrator (mOD) | 575 | ≥ 100 |
| Control Serum (AU/mL) | 19,7 | 19,3-41,3 |

CALIBRATOR

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 3 | 4 | 0 | 0 | 0 | 4 | 6 | 1 | 1 | 7 | 3 | 3 | 5 | 8 | 4 | 2 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature MB

340004611733584200
tTg-G
CALIBRATOR 0.175 mL IVD
LOT 004
2025-01 2°C-8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 3 | 4 | 0 | 0 | 0 | 4 | 6 | 1 | 2 | 2 | 3 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature MB

340004612230000001
tTg-G
CONTROL I+ 0.425 mL IVD
LOT 004
2025-01 2°C-8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. [Signature]

DATE 08.03.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.