



DIESSSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

**QUALITY CONTROL CERTIFICATE**

**KIT NAME 88000-F64 CRESS**

Code	88000-F64		
Lot No.	002	2025-01	2024-03

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D6841	88000-F64 CRESS	002	2025-01	2X6
B6840	CALIBRATORE 88000-F64 CRESS	002	2025-02	1
C6840	CONTROL SERUM 88000-F64 CRESS	002	2025-02	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	284	≥ 100
Control Serum (KU/L)	8,2	5,8-14,3

**CALIBRATOR**

7	0	9	0	0	2	6	5	2	1	6	0	2	6	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES  Operator Signature FB



**CONTROL SERUM**

7	0	9	0	0	2	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB



*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. [Signature]

DATE 12/03/2024

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.