



QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F15 BEAN

Code	88000-F15		
Lot No.	002	2024-10	2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D0491	88000-F15 BEAN	002	2024-10	2x6
B0490	CALIBRATORE 88000-F15 BEAN	002	2025-02	1
C0490	CONTROL SERUM 88000-F15 BEAN	002	2025-02	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	284	≥ 100
Control Serum (KU/L)	8,2	5,8-14,3

CALIBRATOR

6	6	9	0	0	2	5	5	2	1	6	0	2	6	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

669002552160268100
Bean
CALIBRATOR 0.250 mL IVD
LOT 002
2025-02 2°C 8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

6	6	9	0	0	2	5	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

669002527500000001
Bean
CONTROL+ 0.450 mL IVD
LOT 002
2025-02 2°C 8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents Amo

DATE 13.03.24

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.