

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** Q FEVER PHASE II IgM - CHORUS

Code	81172		
Lot No.	017	2025-03	2024-03

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D2090	Q FEVER PHASE II IgM	017	2025-03	6X6
B2090	CALIBRATORE Q FEVER PHASE II IgM	017	2025-03	1
C2090	CONTROL SERUM Q FEVER PHASE II IgM	017	2025-03	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	350	≥ 100
Control Serum (Index)	3,2	2,1-3,8

**CALIBRATOR**

9	4	3	0	1	7	6	0	0	0	0	0	3	4	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB

94301760000340000  
Q Fever phase II IgM  
CALIBRATOR 0.175 mL IVD  
LOT 017  
2025-03  
DIESSE Monteriggioni Italy CE

**CONTROL SERUM**

9	4	3	0	1	7	6	2	1	3	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB

943017621380000001  
Q Fever phase II IgM  
CONTROL+ 0.425 mL IVD  
LOT 017  
2025-03  
DIESSE Monteriggioni Italy CE

**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents Blue

DATE 13/03/2024

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.