

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F76 ALPHA-LACTALBULMIN

Code	88000-F76		
Lot No.	002	2025-01	2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D9861	88000-F76 ALPHA-LACTALBULMIN	002	2025-01	2X6
B9860	CALIBRATORE 88000-F76 ALPHA-LACTALBULMIN	002	2025-02	1
C9860	CONTROL SERUM 88000-F76 ALPHA-LACTALBULMIN	002	2025-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	284	≥ 100
Control Serum (KU/L)	8,2	5,8-14,3

CALIBRATOR

6	1	9	0	0	2	6	5	2	1	6	0	2	6	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *SS*

619002652160268100
Alpha-lactalbumin
CALIBRATOR 0.250 mL IVD
LOT 002
2025-02 2°C 8°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

6	1	9	0	0	2	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *SS*

619002627500000001
Alpha-lactalbumin
CONTROL 0.450 mL IVD
LOT 002
2025-02 2°C 8°C
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Busc*

DATE *18.03.2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.