



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME tTg-A - CHORUS

Code	86058		
Lot No.	010	2025-02	2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D3431	tTg-A	010	2025-02	6X6
B3431	CALIBRATORE tTg-A	010	2025-02	1
C3431	CONTROL SERUM tTg-A	010	2025-02	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	546	≥ 100
Control Serum (AU/mL)	40,6	27,9-54,4

CALIBRATOR

3	3	9	0	1	0	6	2	1	7	3	3	4	5	4	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

339010621733454900
tTg-A
CALIBRATOR 0.175 mL [IVD]
LOT 010
2025-02 2°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

3	3	9	0	1	0	6	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

339010616300000001
tTg-A
CONTROL+ 0.425 mL [IVD]
LOT 010
2025-02 2°C
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE 25/03/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.