

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-FI05 CHOCOLATE

Code	88000-F105		
Lot No.	002	2025-01	2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D1141	88000-FI05 CHOCOLATE	002	2025-01	2X6
B1140	CALIBRATORE 88000-FI05 CHOCOLATE	002	2025-02	1
C1140	CONTROL SERUM 88000-FI05 CHOCOLATE	002	2025-02	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	284	≥ 100
Control Serum (KU/L)	8,2	5,8-14,3

CALIBRATOR

6	4	2	0	0	2	6	5	2	1	6	0	2	6	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature *SB*



CONTROL SERUM

6	4	2	0	0	2	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *SB*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Leone*

DATE *25-03-2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.