

QUALITY CONTROL CERTIFICATE

KIT NAME MUMPS IgG

Code	81075		
Lot No.	007	2025-05	2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D2451	MUMPS IgG	007	2025-05	6X6
B2452	CALIBRATORE MUMPS IgG	007	2025-06	1
C2451	CONTROL SERUM MUMPS IgG	007	2025-06	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	404	≥ 100
Control Serum (Index)	3,0	2,3-4,3
Negative Control (Index)	0,0	< 0,7

CALIBRATOR

0	1	0	0	0	7	7	0	0	0	0	0	3	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB



CONTROL SERUM

0	1	0	0	0	7	7	2	3	4	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 26/03/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.