

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F75 EGG YOLK

Code 88000-F75
Lot No. 002 2025-01 2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D9991	88000-F75 EGG YOLK	002	2025-01	2X6
B9990	CALIBRATORE 88000-F75 EGG YOLK	002	2025-02	1
C9990	CONTROL SERUM 88000-F75 EGG YOLK	002	2025-02	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	284	≥ 100
Control Serum (KU/L)	8,2	5,8-14,3

CALIBRATOR

6	2	0	0	0	2	6	5	2	1	6	0	2	6	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature



CONTROL SERUM

6	2	0	0	0	2	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 27-03-2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.