



QUALITY CONTROL CERTIFICATE

KIT NAME CARDIOLIPIN-M - CHORUS

Code	86048		
Lot No.	004	2025-01	2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D3331	CARDIOLIPIN-M	004	2025-01	6X6
B3331	CALIBRATORE CARDIOLIPIN-M	004	2025-01	1
C3331	CONTROL SERUM CARDIOLIPIN-M	004	2025-01	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	570	≥ 100
Control Serum (MPLU/mL)	58,9	33,3-100,0

CALIBRATOR

3	3	6	0	0	4	6	5	1	8	2	7	3	8	3	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

336004651627383500
Cardiolipin-M
CALIBRATOR 0.175 mL **IVD**
LOT 004
2025-01 2°C-8°C
DIESSE Monteriggioni Italy **CE**
B3331

CONTROL SERUM

3	3	6	0	0	4	6	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

336004619350000001
Cardiolipin-M
CONTROL+ 0.425 mL **IVD**
LOT 004
2025-01 2°C-8°C
DIESSE Monteriggioni Italy **CE**
C3331

*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. [Signature]

DATE 03/04/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.