



DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

QUALITY CONTROL CERTIFICATE

KIT NAME SANITIZING SOLUTION

Code	83604		
Lot No.	075	2027-03	2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
B2563	SANITIZING SOLUTION	075	2027-03	1

CONTROLS

	Found results	Expected results
Density	1.060	1.060-1.100
pH	13,45	12-14

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 04-04-2024