

DIESE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

**PRODUCT NAME:** VACU-TEC

<b>Code</b>	10602
<b>Lot No.</b>	4057  2024-12  2024-02
<b>Reference</b>	IO-09-88

**ON LINE QUALITY CONTROLS** (model CDP not enclosed)

**QC Passed**

- sodium citrate solution concentration
- sodium citrate volume
- vacuum level (\*)
- correct placement of the cap
- no extra body inside the product
- (\*) Only for VACU-CODE and VACU-TEC tubes

**PACKAGING QUALITY CONTROL** (model CDP not enclosed)

**QC Passed**

- control of the number of the product for each pack

**QUALITY CONTROLS ON FINISHED PRODUCT** (model CDP not enclosed)

**QC Passed**

- pack integrity
- data printed on the product label

**QC Passed**

N/A

**Irradiation treatment Certificate n° 2025/2024 date 28/03/2024**

**QC Passed**

N/A

**Gamma rays treatment process in compliance with the specifications of the supply Agreement**

Document number:

ODV N.A.

***This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.***

**APPROVED** RESPONSIBLE HEAD, Quality Control Reagents..... 

**DATE** ...04/04/2024