



DIESSSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F13 PEANUT

Code	88000-F13		
Lot No.	002	2025-01	2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D9941	88000-F13 PEANUT	002	2025-01	2x6
B9940	CALIBRATORE 88000-F13 PEANUT	002	2025-02	1
C9940	CONTROL SERUM 88000-F13 PEANUT	002	2025-02	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	284	≥ 100
Control Serum (KU/L)	8,2	5,8-14,3

CALIBRATOR

6	1	7	0	0	2	6	5	2	1	6	0	2	6	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*

617002652160268100  
Peanut  
CALIBRATOR 0,250 mL IVD  
LOT 002  
 2025-02 2°C-8°C  
DIESSSE Monteriggioni Italy CE  
B9940

CONTROL SERUM

6	1	7	0	0	2	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*

617002627500000001  
Peanut  
CONTROL 0,450 mL IVD  
LOT 002  
 2025-02 2°C-8°C  
DIESSSE Monteriggioni Italy CE  
C9940

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE *04.06.2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.