



QUALITY CONTROL CERTIFICATE

KIT NAME 88000-K82 LATEX

Code	88000-K82		
Lot No.	002	2025-01	2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D6941	88000-K82 LATEX	002	2025-01	2X6
B6940	CALIBRATORE 88000-K82 LATEX	002	2025-02	1
C6940	CONTROL SERUM 88000-K82 LATEX	002	2025-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	284	≥ 100
Control Serum (KU/L)	8,2	5,8-14,3

CALIBRATOR

6	5	3	0	0	2	6	5	2	1	6	0	2	6	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

653002652160268100
Latex
CALIBRATOR 0.250 mL IVD
LOT 002
2025-02 2°C 8°C B6940
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

6	5	3	0	0	2	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

653002627500000001
Latex
CONTROL+ 0.450 mL IVD
LOT 002
2025-02 2°C 8°C C6940
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... [Signature]

DATE 04/04/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.