

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME tTg-A - CHORUS

Code 86058

Lot No. 011

2025-02

2024-02

REAGENTS

Label	Description	Lot	Exp.	NO
D3431	tTg-A	011	2025-02	6X6
B3431	CALIBRATORE tTg-A	011	2025-02	1
C3431	CONTROL SERUM tTg-A	011	2025-02	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	481	≥ 100
Control Serum (AU/mL)	27,3	25,5-55,4

CALIBRATOR

3	3	9	0	1	1	6	1	2	0	3	3	4	2	4	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *[Signature]*



CONTROL SERUM

3	3	9	0	1	1	6	1	7	3	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *[Signature]*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 08.06.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.