

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME ANTI-CCP - CHORUS

Code 86094

Lot No. 019

2025-03

2024-03

## REAGENTS

Label	Description	Lot.	Exp.	NO
D8671	ANTI-CCP	019	2025-03	6x6
B8671	CALIBRATORE ANTI-CCP	019	2025-03	1
C8671	CONTROL SERUM ANTI-CCP	019	2025-03	1

## CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	488	$\geq 100$
Control Serum (AU/mL)	54,5	36,5-78,9

## CALIBRATOR

2	5	6	0	1	9	6	3	2	3	4	2	3	8	6	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *MB*




## CONTROL SERUM

2	5	6	0	1	9	6	1	8	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *MB*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE *16.06.2024*

## WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.