

QUALITY CONTROL CERTIFICATE

KIT NAME CALPROTECTIN 3RD GENERATION- CHORUS

Code	86603			
Lot No.	016	2025-03	2024-03	

REAGENTS

Label	Description	Lot.	Exp.	NO
D8430	CALPROTECTIN 3 RD GENERATION	016	2025-03	6X6
B8430	CALIBRATORE CALPROTECTIN 3 RD GENERATION	016	2025-03	1
C8430	CONTROL CALPROTECTIN 3 RD GENERATION	016	2025-03	1
G8430	SAMPLE DILUENT/NEGATIVE CONTROL CALPROTECTIN 3 RD GENERATION	016	2025-09	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	380	≥ 100
Control (mg/Kg)	551,9	305,8-766,7
sample dil/neg control Calp 3 RD (mg/Kg)	<24,0	< 51,0

CALIBRATOR

9	9	3	0	1	6	6	3	2	2	3	5	3	5	4	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature



CONTROL SERUM

9	9	3	0	1	6	6	2	0	3	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED BY RESPONSIBLE HEAD, Quality Control Reagents.

DATE 16.04.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.