

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME BRUCELLA IgG CHORUS

Code	81164		
Lot No.	024	2025-04	2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D4110	Brucella IgG	024	2025-04	6X6
B4110	CALIBRATORE Brucella IgG	024	2025-04	1
C4110	CONTROL SERUM Brucella IgG	024	2025-04	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	577	≥ 100
Control (index)	1,7	1,3-2,4

CALIBRATOR

1	4	6	0	2	4	6	0	0	0	0	0	5	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



CONTROL SERUM

1	4	6	0	2	4	6	1	3	2	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 12/04/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.