

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME HELICOBACTER PYLORI IgG -CHORUS

Code	81060		
Lot No.	024	2025-04	2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2392	Helicobacter Pylori IgG	024	2025-04	6X6
B2393	CALIBRATORE Helicobacter Pylori IgG	024	2025-04	1
C2392	CONTROL SERUM Helicobacter Pylori IgG	024	2025-04	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	320	≥ 100
Control Serum (UA/mL)	73,2	59,4-200,0
Negative Control (UA/mL)	<5,0	< 5

CALIBRATOR

0	0	8	0	2	4	6	1	2	0	3	6	3	1	4	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *[Signature]*



CONTROL SERUM

0	0	8	0	2	4	6	2	3	4	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *[Signature]*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 22-04-2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.