

DIESSE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
 Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
 20144 Milano (Italy) Via A. Solari 19 sc. 6

PRODUCT NAME: VES-TEC

Code	10214		
Lot No.	1624	2025-06	2024-04
Reference IO-09-89			

ON LINE QUALITY CONTROLS (model CDP not enclosed)

QC Passed

- sodium citrate solution concentration
- sodium citrate volume
- vacuum level (*)
- correct placement of the cap
- no extra body inside the product
- (*) Only for VACU-CODE and VACU-TEC tubes

PACKAGING QUALITY CONTROL (model CDP not enclosed)

QC Passed

- control of the number of the product for each pack

QUALITY CONTROLS ON FINISHED PRODUCT (model CDP not enclosed)

QC Passed

- pack integrity
- data printed on the product label

QC Passed

N/A

Irradiation treatment Certificate n° date

QC Passed

N/A

Gamma rays treatment process in compliance with the specifications of the supply Agreement

Document number:

ODV N.A.

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 22 di 04