



QUALITY CONTROL CERTIFICATE

KIT NAME PARVOVIRUS B19 IgG - CHORUS

Code	81150		
Lot No.	024	2025-04	2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D0220	Parvovirus B19 IgG	024	2025-04	6X6
B0220	CALIBRATORE Parvovirus B19 IgG	024	2025-04	1
C0220	CONTROL SERUM Parvovirus B19 IgG	024	2025-04	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	430	≥ 100
Control Serum (Index)	2,8	2,2-4,0

CALIBRATOR

0	1	1	0	2	4	6	0	0	0	0	0	3	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

01102460000330000
Parvovirus B19 IgG
CALIBRATOR 0.175 mL IVD
LOT 024
2025-04 2°C-8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

0	1	1	0	2	4	6	2	2	4	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

01102462240000001
Parvovirus B19 IgG
CONTROL+ 0.425 mL IVD
LOT 024
2025-04 2°C-8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. [Signature]

DATE 30/04/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.