

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TOXOPLASMA IgM - CHORUS

Code 81041

Lot No. 016

2025-03

2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2281	TOXOPLASMA IgM	016	2025-03	6X6
B2281	CALIBRATORE TOXOPLASMA IgM	016	2025-03	1
C2281	CONTROL SERUM TOXOPLASMA IgM	016	2025-03	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	420	≥ 100
Control Serum (INDEX)	2,1	1,5-2,7
Negative Control (INDEX)	0,0	<0,7


CALIBRATOR

1 3 1 0 1 6 6 0 0 0 0 0 4 8 0 0 0 0

Label checked with barcode reader

YES ☒

Operator Signature




CONTROL SERUM

1 3 1 0 1 6 6 1 5 2 7 0 0 0 0 0 0 1

Label checked with barcode reader

YES ☒

Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE

29.04.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.