



DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** DEAMIDATED GLIADIN-G - CHORUS

Code	86100
Lot No.	024  2025-04  2024-04

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D8660	Deamidated Gliadin-G	024	2025-04	6X6
B8660	CALIBRATORE Deamidated Gliadin-G	024	2025-04	1
C8660	CONTROL SERUM Deamidated Gliadin-G	024	2025-04	1


**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	610	≥ 100
Control Serum (AU/mL)	45,1	25,5-64,3

**CALIBRATOR**

2	5	5	0	2	4	6	1	2	0	3	4	5	4	5	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature



**CONTROL SERUM**

2	5	5	0	2	4	6	1	3	2	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature



**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents   
DATE 26. di. 2024

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.