DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME TOXOPLASMA IgG AVIDITY - CHORUS

Code **81098**

Lot No. 024 2025-04 2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2531	TOXOPLASMA IgG AVIDITY	024	2025-04	2x6
C2531	CONTROL SERUM HIGH TOXOPLASMA IgG AVIDITY	024	2025-04	1
G2531	SAMPLE DILUENT	024	2025-10	1

CONTROLS

	Found results	Expected results
Control Serum (%)	71,04%	> 50 ≤ 90 %

CALIBRATOR

Label checked with barcode reader YES ☐ Operator Signature .....

STICK LABEL

CONTROL SERUM

0	8	3	0	2	4	6	5	0	9	0	0	0	0	0	0	0	1

Label checked with barcode reader YES ☒ Operator Signature FB

083024650900000001
Toxoplasma IgG Avidity
CONTROL HIGH 0.3 mL IVD
LOT 024
2025-04
2-8°C
C2531
DIESSE Monteriggioni Italy CE 0123

This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 30/04/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma.All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.