



DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** 88000-F12 PEA

Code	88000-F12		
Lot No.	024	2025-04	2024-04

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D0481	88000-F12 PEA	024	2025-04	2x6
B0480	CALIBRATORE 88000-F12 PEA	024	2025-06	1
C0480	CONTROL SERUM 88000-F12 PEA	024	2025-06	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	344	≥ 100
Control Serum (KU/L)	11,6	7,4-16,1

**CALIBRATOR**

6	7	4	0	2	4	6	5	1	7	5	5	2	6	7	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB

674024651755267100  
Pea  
CALIBRATOR 0.250 mL IVD  
LOT 024  
2025-06 2°C 8°C  
DIESSE Monteriggioni Italy CE

**CONTROL SERUM**

6	7	4	0	2	4	6	2	7	5	0	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB

674024627500000001  
Pea  
CONTROL+ 0.450 mL IVD  
LOT 024  
2025-06 2°C 8°C  
DIESSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... FB

DATE 02/05/2024

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**  
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.