

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CALPROTECTIN K - CHORUS

Code 86601

Lot No. 009



2025-02



2024-02

REAGENTS

Label	Description	Lot	Exp.	NO
D9970	CALPROTECTIN	009	2025-02	2X6
B9970	CALIBRATORE CALPROTECTIN	009	2025-02	1
C9970	CONTROL CALPROTECTIN	009	2025-02	1
G9502	SAMPLE DILUENT	009	2026-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	340	≥ 100
Control Serum (mg/Kg)	625,5	376,2-1000,0

CALIBRATOR

9	2	9	0	0	9	6	4	2	4	4	5	2	5	5	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

MB



CONTROL SERUM

9	2	9	0	0	9	6	2	8	5	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

MB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

MB

DATE 29.06.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.