

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-D1 DERMATOPHAGOIDES PTERONYSSINUS

Code	88000-D1		
Lot No.	024	2025-04	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D9791	88000-D1 DERMATOPHAGOIDES PTERONYSSINUS	024	2025-04	2x6
B9790	CALIBRATORE 88000-D1 DERMATOPHAGOIDES PTERONYSSINUS	024	2025-06	1
C9790	CONTROL SERUM 88000-D1 DERMATOPHAGOIDES PTERONYSSINUS	024	2025-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	344	≥ 100
Control Serum (KU/L)	11,6	7,4-16,1

CALIBRATOR

6 0 2 0 2 4 6 5 1 7 5 5 2 6 7 1 0 0

Label checked with barcode reader YES Operator Signature

602024651755267100
Dermatophagoides pteronyssinus
CALIBRATOR 0,250 mL IVD
LOT 024
2025-06 2°C-8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

6 0 2 0 2 4 6 2 7 5 0 0 0 0 0 0 0 1

Label checked with barcode reader YES Operator Signature

602024627500000001
Dermatophagoides pteronyssinus
CONTROL+ 0,450 mL IVD
LOT 024
2025-06 2°C-8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 02/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.