



DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F8 MAIZE

Code	88000- F8		
Lot No.	024	2025-04	2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D0511	88000-F8 MAIZE	024	2025-04	2x6
B0510	CALIBRATORE 88000-F8 MAIZE	024	2025-06	1
C0510	CONTROL SERUM 88000-F8 MAIZE	024	2025-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	344	≥ 100
Control Serum (KU/L)	11,6	7,4-16,1

CALIBRATOR

6 6 6 0 2 4 6 5 1 7 5 5 2 6 7 1 0 0

Label checked with barcode reader YES  Operator Signature *FB*

666024651755267100  
Maize  
CALIBRATOR 0.250 mL [IVD]  
LOT 024  
2025-06 2°C 8°C  
DIESSE Monteriggioni Italy CE

CONTROL SERUM

6 6 6 0 2 4 6 2 7 5 0 0 0 0 0 0 0 1

Label checked with barcode reader YES  Operator Signature *FB*

666024627500000001  
Maize  
CONTROL 0.480 mL [IVD]  
LOT 024  
2025-06 2°C 8°C  
DIESSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Bone*

DATE 23/04/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.