

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F93 CACAO

Code	88000-F93		
Lot No.	024	2025-04	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D9881	88000-F93 CACAO	024	2025-04	2X6
B9880	CALIBRATORE 88000-F93 CACAO	024	2025-06	1
C9880	CONTROL SERUM 88000-F93 CACAO	024	2025-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	344	≥ 100
Control Serum (KU/L)	11,6	7,4-16,1

CALIBRATOR

6	3	3	0	2	4	6	5	1	7	5	5	2	6	7	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

633024651755267100
Cacao
CALIBRATOR 0.250 mL [IVD]
LOT 024
2025-06 2°C-8°C
DIESE Monteriggioni Italy CE

CONTROL SERUM

6	3	3	0	2	4	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

633024627500000001
Cacao
CONTROL 0.450 mL [IVD]
LOT 024
2025-06 2°C-8°C
DIESE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE *07/05/2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.