

DIESSA DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME FSH

| | |
|---------|---------|
| Code | 86510 |
| Lot No. | 024 |
| | 2025-01 |
| | 2024-04 |

REAGENTS

| Label | Description | Lot. | Exp. | NO |
|-------|-------------------|------|---------|-----|
| D3230 | FSH | 024 | 2025-01 | 6X6 |
| B3230 | CALIBRATOR FSH | 024 | 2025-01 | 1 |
| C3230 | CONTROL SERUM FSH | 024 | 2025-01 | 1 |

CONTROLS

| | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

| | | |
|------------------------|------|----------|
| Calibrator (mOD) | 290 | ≥ 100 |
| Control Serum (mIU/ml) | 11,6 | 6,5-15,0 |

CALIBRATOR

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 9 | 2 | 0 | 0 | 2 | 4 | 5 | 5 | 1 | 8 | 3 | 6 | 3 | 7 | 5 | 0 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature *FB*

920024551836375000
FSH
CALIBRATOR 0,600 mL IVD
LOT 024
2025-01
2°C-8°C
DIESSA Monteriggioni Italy CE

CONTROL SERUM

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 9 | 2 | 0 | 0 | 2 | 4 | 5 | 1 | 9 | 3 | 6 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature *FB*

920024519360000001
FSH
CONTROL+ 1 mL IVD
LOT 024
2025-01
2°C-8°C
DIESSA Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE *07-05-2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.