

QUALITY CONTROL CERTIFICATE

KIT NAME LKM-1 CHORUS

Code	86080			
Lot No.	017	2025-03	2024-03	

REAGENTS

Label	Description	Lot.	Exp.	NO
D3641	LKM-1	017	2025-03	6X6
B3641	CALIBRATORE LKM-1	017	2025-03	1
C3641	CONTROL SERUM LKM-1	017	2025-03	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	510	≥ 100
Control Serum (AU/mL)	58,9	30,6-75,0

CALIBRATOR

7	0	3	0	1	7	6	4	1	9	3	3	5	0	5	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

703017641933505900
LKM-1
CALIBRATOR 0.175 mL IVD
LOT 017
 2025-03 2°C-8°C
DIESSSE Monteriggioni Italy CE
B0641

CONTROL SERUM

7	0	3	0	1	7	6	1	4	2	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

703017614260000001
LKM-1
CONTROL 0.425 mL IVD
LOT 017
 2025-03 2°C-8°C
DIESSSE Monteriggioni Italy CE
C3641

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 07/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.