



QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F49 APPLE

Code	88000-F49		
Lot No.	024	2025-04	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D1061	88000-F49 APPLE	024	2025-04	2X6
B1060	CALIBRATORE 88000-F49 APPLE	024	2025-06	1
C1060	CONTROL SERUM 88000-F49 APPLE	024	2025-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	344	≥ 100
Control Serum (KU/L)	11,6	7,4-16,1

CALIBRATOR

6	6	0	0	2	4	6	5	1	7	5	5	2	6	7	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

660024651755267100
Apple
CALIBRATOR 0,250 mL IVD
LOT 024
2025-06 B1060
DIESSE Monteriggioni Italy

CONTROL SERUM

6	6	0	0	2	4	6	2	7	5	0	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

660024627500000001
Apple
CONTROL+ 0,450 mL IVD
LOT 024
2025-06 C1060
DIESSE Monteriggioni Italy

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents Beuca

DATE 07/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.