



QUALITY CONTROL CERTIFICATE

KIT NAME tTg-G - CHORUS

Code	86060		
Lot No.	018-A	2025-03	2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D3451	tTg-G	018	2025-03	6X6
B3451	CALIBRATORE tTg-G	018-1	2025-03	1
C3451	CONTROL SERUM tTg-G	018-1	2025-03	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	420	≥ 100
Control Serum (AU/mL)	49,1	40,0-94,3

CALIBRATOR

3	4	0	0	1	8	6	2	1	5	3	0	3	6	3	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES  Operator Signature FB



CONTROL SERUM

3	4	0	0	1	8	6	2	0	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB



*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents Blue

DATE 06/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.