

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME ANTI-CCP - CHORUS

Code 86094

Lot No. 020

2025-03

2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D8671	ANTI-CCP	020	2025-03	6x6
B8671	CALIBRATORE ANTI-CCP	020	2025-03	1
C8671	CONTROL SERUM ANTI-CCP	020	2025-03	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	374	≥ 100
Control Serum (AU/mL)	48,5	41,9-86,8


CALIBRATOR

2 5 6 0 2 0 6 3 2 3 4 2 3 5 6 5 0 0

Label checked with barcode reader

YES ☒

Operator Signature *FB*



CONTROL SERUM

2 5 6 0 2 0 6 2 0 3 7 0 0 0 0 0 0 1

Label checked with barcode reader

YES ☒

Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *AS*

DATE 08/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.