

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CYTOMEGALOVIRUS IgG -CHORUS

Code 81010

Lot No. 017

2025-03

2024-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D2161	Cytomegalovirus IgG	017	2025-03	6x6
B2162	CALIBRATORE Cytomegalovirus IgG	017	2025-03	1
C2161	CONTROL SERUM Cytomegalovirus IgG	017	2025-03	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	264	≥ 100
Control Serum (IU/mL)	6,3	4,3-12,4
Negative Control (IU/mL)	<0,5	<0,5

CALIBRATOR

0 0 1 0 1 7 6 4 2 6 3 7 3 1 5 8 0 0

Label checked with barcode reader

YES ☒

Operator Signature *MB*




CONTROL SERUM

0 0 1 0 1 7 6 2 3 4 2 0 0 0 0 0 0 1

Label checked with barcode reader

YES ☒

Operator Signature *MB*



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *MB*

DATE 08.05.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.