

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TREPONEMA IgG - CHORUS

Code 81050/12

Lot No. 024



2025-04



2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2321	TREPONEMA IgG	024	2025-04	2X6
B2322	CALIBRATORE TREPONEMA IgG	024	2025-04	1
C2321	CONTROL SERUM TREPONEMA IgG	024	2025-04	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	480	≥ 100
Control Serum (INDEX)	3,0	2,0-3,8
Negative Control (INDEX)	0,0	$<0,7$

CALIBRATOR

0	1	2	0	2	4	6	0	0	0	0	0	3	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature MB



CONTROL SERUM

0	1	2	0	2	4	6	2	0	3	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature MB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 09/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.