



QUALITY CONTROL CERTIFICATE

KIT NAME 88000- E5- DOG DANDER

Code	88000-E5		
Lot No.	024	2025-04	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D6911	88000- E5 DOG DANDER	024	2025-04	2x6
B6910	CALIBRATORE 88000- E5 DOG DANDER	024	2025-06	1
C6910	CONTROL SERUM 88000- E5 DOG DANDER	024	2025-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	344	≥ 100
Control Serum (KU/L)	11,6	7,4-16,1

CALIBRATOR

6	4	8	0	2	4	6	5	1	7	5	5	2	6	7	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

648024651755267100
Dog dander
CALIBRATOR 0.250 mL [IVD]
LOT 024
 2025-06 2°C B6910
DIESSE Monteriggioni Italy

CONTROL SERUM

6	4	8	0	2	4	6	2	7	5	0	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

648024627500000001
Dog dander
CONTROL 0.460 mL [IVD]
LOT 024
 2025-06 2°C C6910
DIESSE Monteriggioni Italy

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *[Signature]*
DATE 10/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.