

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME RUBELLA IgG - CHORUS

Code	81030		
Lot No.	033	2025-05	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D2231	RUBELLA IgG	033	2025-05	6X6
B2232	CALIBRATORE RUBELLA IgG	033	2025-05	1
C2231	CONTROL SERUM RUBELLA IgG	033	2025-05	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	403	≥ 100
Control Serum (IU/mL)	30,5	21,3-50,0
Negative Control (IU/mL)	<5,0	< 5

CALIBRATOR

0	8	1	0	3	3	6	5	1	8	3	0	4	3	4	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

0	8	1	0	3	3	6	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 15/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.