

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME RUBELLA IgM - CHORUS

Code	81031		
Lot No.	024	2025-04	2024-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2241	RUBELLA IgM	024	2025-04	6X6
B2241	CALIBRATORE RUBELLA IgM	024	2025-04	1
C2241	CONTROL SERUM	024	2025-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	505	≥ 100
Control Serum (INDEX)	2,2	1,4-2,7
Negative Control (INDEX)	0,0	< 0,7

CALIBRATOR

1	3	5	0	2	4	6	0	0	0	0	0	4	9	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

13502460000490000
Rubella IgM
CALIBRATOR 0.175 mL IVD
LOT 024
2025-04
2°C - 8°C
DIESSE Monteriggioni Italy CE 0123 B2241

CONTROL SERUM

1	3	5	0	2	4	6	1	4	2	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

135024614270000001
Rubella IgM
CONTROL 0.425 mL IVD
LOT 024
2025-04
2°C - 8°C
DIESSE Monteriggioni Italy CE 0123 C2241

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 15.05.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.