

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-H2 HOUSE DUST HOLLISTER-STIER LABS

Code	88000-H2		
Lot No.	030	2025-04	2024-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D3361	88000-H2 HOUSE DUST HOLLISTER-STIER LABS	030	2025-04	2X6
B3360	CALIBRATORE 88000-H2 HOUSE DUST HOLLISTER-STIER LABS	030	2025-06	1
C3360	CONTROL SERUM 88000-H2 HOUSE DUST HOLLISTER-STIER LABS	030	2025-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	300	≥ 100
Control Serum (KU/L)	10,1	7,4-16,1

CALIBRATOR

6	4	1	0	3	0	6	5	1	7	5	5	2	6	7	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

641030651755267100
House dust
Hollister-stier labs
CALIBRATOR 0.250 mL IVD
LOT 030
2025-06
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

6	4	1	0	3	0	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

641030627500000001
House dust
Hollister-stier labs
CONTROL 0.480 mL IVD
LOT 030
2025-06
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 17/05/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.